

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

CB TEST CERTIFICATE

Product	Medical Power Supply
Name and address of the applicant	SL POWER ELECTRONICS CORP 6050 KING DRIVE, BLDG A VENTURA, CA 93003 UNITED STATES
Name and address of the manufacturer	SL POWER ELECTRONICS CORP 6050 KING DRIVE, BLDG A VENTURA, CA 93003 UNITED STATES
Name and address of the factory	INDUSTRIAS S L S A DE C V CIRCUITO SIGLO XXI 2055 COL PARQUE INDUSTRIAL EX-XXI, 21254 MEXICALI BC MEXICO
Note: <i>When more than one factory, please report on page 2</i>	<input checked="" type="checkbox"/> Additional Information on page 2
Ratings and principal characteristics	See Page 2
Trademark (if any)	
Type of Customer's Testing Facility (CTF) Stage used	CTF Stage 3
Model / Type Ref.	SE240AXXYZWW See Page 2
Additional information (if necessary may also be reported on page 2)	The report was revised to include technical modifications. <input checked="" type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1-6:2010/AMD1:2013, IEC 60601-1-6:2010, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005
As shown in the Test Report Ref. No. which forms part of this Certificate	E116994-D1021-1/A1/C0-CB issued on 2019-04-29

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2019-05-10

Original Issue Date: 2018-11-08

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

US-32729-M1-UL

Model Details:

SE240AXXYZWW where S can be M or T, XX represents the output voltage from 12V to 48V, Y represents any number from 0 to 9 or blank, Z represents F for class I construction, W represents any number from 0 to 9 or blank, unless specified below.

Model ME240A24YYF02, ME240B2400F02

Factories:

SL XIANGHE POWER ELECTRONICS CORP

No.B-02-03,North side of Landscape Ave, Qibu District,Environmental Industrial Park Xianghe, Hebei 065400 CHINA

Ratings:

Input: 100-240V~, 50-60Hz, 2.8A

Output: See Model Differences in Test Report for the output ratings

Model ME240A1251Z01:

Input: 100-110Vac, 50-60Hz, 2.8A; Output: 12Vdc, 16.6A

Input: 110-240Vac, 50-60Hz, 2.8A; Output: 12Vdc, 16.6A

Model ME240A2451Z01:

Input: 100-110Vac, 50-60Hz, 2.8A; Output: 24Vdc, 9.17A

Input: 110-240Vac, 50-60Hz, 2.8A; Output: 24Vdc, 10.0A

Model ME240A2851Z01:

Input: 100-110Vac, 50-60Hz, 2.8A; Output: 28Vdc, 7.86A

Input: 110-240Vac, 50-60Hz, 2.8A; Output: 28Vdc, 8.60A

Model ME240A4851Z01:

Input: 100-110Vac, 50-60Hz, 2.8A; Output: 48Vdc, 5.0A

Input: 110-240Vac, 50-60Hz, 2.8A; Output: 48Vdc, 5.0A

Model ME240A2451F02:

Input: 100-110Vac, 50-60Hz, 2.8A; Output: 25Vdc, 8.80A

Input: 110-240Vac, 50-60Hz, 2.8A; Output: 25Vdc, 9.60A

Additional Information:

Additionally evaluated to EN 60601 1:2006/ A1:2013/ A12:2014 and EN 60601-1-6:2010/ A1:2015; National Differences specified in the CB Test Report.

The original report was modified to include the following changes/additions:

Addition of a new model and alternate component.

See Test Report for details.

Additional information (if necessary)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2019-05-10

Original Issue Date: 2018-11-08

Signature:

Jolanta M. Wroblewska