

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Medical Power Supply
Name and address of the applicant	BRIDGEPOWER CORP (GOSAEK-DONG) 16 OMOKCHEN-RO 132BEON-GIL GWONSEON-GU SUWON-SI GYEONGGI 441-813 KOREA
Name and address of the manufacturer	BRIDGEPOWER CORP (GOSAEK-DONG) 16 OMOKCHEN-RO 132BEON-GIL GWONSEON-GU SUWON-SI GYEONGGI 441-813 KOREA
Name and address of the factory <i>Note: When more than one factory, please report on page 2</i>	BRIDGEPOWER CORP (GOSAEK-DONG) 16 OMOKCHEN-RO 132BEON-GIL GWONSEON-GU SUWON-SI GYEONGGI 441-813 KOREA <input checked="" type="checkbox"/> Additional Information on page 2
Ratings and principal characteristics	Rated Input: 100-240 Vac, 50-60 Hz, 0.7 A (0.7A - 0.4A) Rated Output: - Output 1(Main) : 5 Vdc, 2.4 A - Output 2(USB) : 5 Vdc, 2.4 A
Trademark (if any)	None
Type of Customer's Testing Facility (CTF) Stage used	CTF Stage 2
Model / Type Ref.	(1)E30D(2)(3)(4)(5), B(1)030D(2)(4) See Page 2
Additional information (if necessary may also be reported on page 2)	National differences specified in the CB Test Report. <input type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1-6:2010/AMD1:2013, IEC 60601-1-6:2010, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005
As shown in the Test Report Ref. No. which forms part of this Certificate	E302267-D1043-1/A0/C0-CB issued on 2019-08-27


This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

Date: 2019-09-06

Signature:



Jan-Erik Storgaard

For full legal entity names see www.ul.com/ncbnames



Ref. Certif. No.

DK-87464-UL

Model Details:

(1)E30D(2)(3)(4)(5) and B(1)030D(2)(4) (Where (1), (2), (3), (4), (5) may alphanumeric, "For marketing purpose and no impact safety related critical components and constructions")

- (1) Family Related Designs: A to Z
- (2) Output: 05
- (3) Standards Output Cord Options Number: 00 thru 99
- (4) Standard Input Connector Options : F - Class I appliance inlet type (IEC320-C14)
- (5) Model Configuration Number: 00 thru 99

Factories:

WENDENG JEIL ELECTRONICS CO LTD
XIAMEN ROAD NO.2 WENDENG ECONOMIC DEVELOPMENT ZONE WEIHAI CITY SHANDONG PROVINCE
CHINA

BRIDGEPOWER VINA CO LTD
LOT B9 THUY VAN INDUSTRIAL ZONE VIET TRI CITY PHU THO PROVINCE
VIETNAM

Additional information (if necessary)



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2019-09-06

Signature: 
Jan-Erik Storgaard