

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEME****CB TEST CERTIFICATE**

Product	Power supply
Name and address of the applicant	SL POWER ELECTRONICS CORP BLDG A 6050 KING DR VENTURA, CA 93003 USA
Name and address of the manufacturer	SL POWER ELECTRONICS CORP BLDG A 6050 KING DR VENTURA, CA 93003 USA
Name and address of the factory <i>Note: When more than one factory, please report on page 2</i>	SL XIANGHE POWER ELECTRONICS CORP No.B-02-03, North side of Landscape Ave, Qibu Distr Environmental Industrial Park Xianghe, Hebei 065400 China <input checked="" type="checkbox"/> <a href="#">Additional Information on page 2</a>
Ratings and principal characteristics	See Page 2
Trademark / Brand (if any)	
Type of Customer's Testing Facility (CTF) Stage used	CTF Stage 3
Model / Type Ref.	NGB250SXXYZZ See Page 2
Additional information (if necessary may also be reported on page 2)	Additionally evaluated to EN 60601 1:2006/ A1:2013/ A12:2014; National Differences specified in the CB Test Report. <input checked="" type="checkbox"/> <a href="#">Additional Information on page 2</a>
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005
As shown in the Test Report Ref. No. which forms part of this Certificate	E511446-D1000-1/A0/C0-CB issued on 2019-12-17

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2019-12-19

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

**US-35016-UL**

**Model Details:**

Model: NGB250SXXYZZ, Where XX represents the output voltage which may be any number from 12 to 48. Y can be C (for Class II); ZZ can be any number between 00-99, blank or any letter from AA to ZZ, only for market purpose, not influence safety function.

**Factories:**

INDUSTRIAS S L S A DE C V  
CIRCUITO SIGLO XXI 2055 COL PARQUE INDUSTRIAL EX-XXI 21254 MEXICALI BC  
Mexico

**Ratings:**

Input: 100-240 Vac, 50-60 Hz, 3.0A

**Output:**

Model NGB250S12C

For convection: 12Vdc/12.09A

For 400LFM: 12Vdc/19.1A

Model NGB250S15C

For convection: 15Vdc/10.34A

For 400LFM: 15Vdc/15.3A

Model NGB250S24C

For convection: 24Vdc/7.3A

For 400LFM: 24V/10.4A

Model NGB250S48C

For convection: 48Vdc/3.65A

For 400LFM: 48Vdc/5.2A

**Additional Information:**

The risk management requirements of the standard were not addressed.

**Additional information (if necessary)**



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2019-12-19

Signature:

Jolanta M. Wroblewska