

SG PSB-MD-00116

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product

Switch mode power supplies (Switch Mode Power Supply for Building-in)

Name and address of the applicant

Astec International Ltd.

16th Floor, Lu Plaza, 2 Wing Yip Street, Kwun Tong
Kowloon, HONG KONG

Name and address of the manufacturer

Astec International Ltd.

16th Floor, Lu Plaza, 2 Wing Yip Street, Kwun Tong, Kowloon, HONG KONG

Name and address of the factory

Astec Power Philippines Inc. 104 Laguna Boulevard, Laguna Technopark, 4026 Sta. Rosa Laguna, PHILIPPINES

For further information please see attachment

Ratings and principal characteristics

Rated Input : Rated Output : 100-240V, 1.8A, 50/60Hz See test report for details

Construction:

Built-in

Protection Class:

Built-in component,

consider in end system

Degree of Protection:

: IPX0

Model/type Ref.

NPT42-M, NPT43-M, NPT44-M

Additional information (if necessary)

See Test Report for National Differences and Group Differences
The risk management requirements of the standard were not addressed.

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005

IEC 60601-1:2005/AMD1:2012

as shown in the Test Report Ref. No. which forms part of this certificate

TÜV SÜD PSB Pte Ltd 211-600540-000

This CB Test Certificate is issued by the National Certification Body Ce Certificat d'essai OC est établi par l'Organisme National de Certification

Date,

2017-05-29

CBS 17 05 13890 02850

Ars.

(Yager Bi)



Product Service

TÜV SÜD PSB Pte Ltd · 1 Science Park Drive · Singapore 118221



SG PSB-MD-00116

Additional factory information:

Name and address of the factory Nom et adresse de l'usine Astec Power Philippines Inc. Main Avenue Corner Rd "J", Cavite Export Processing Zone, Rosario, Cavite, Philippines

(Yager Bi)

Jrz.

Signature:



Date,

2017-05-29

CBS 17 05 13890 02850